



## **Test Update 787**

**Posted Date** 12/08/2021  
**Effective Date** 11/18/2021

### **RECALL**

#### **ACD-A solution vacutainer**

**Item#:** 364606

**Manufacturer:** BD

**Lot#:** 0316293

**Expiration:** 11/30/2022

Effective November 18, 2021, MLabs is passing along recall information from the manufacturer Becton Dickinson (BD) that ACD-A solution vacutainers are being recalled for lot 0316293 expiration: 11/30/2022. The manufacturer is asking that these vacutainers be discarded. These tubes may underfill or exhibit draw volumes below labeled product claims. Underfilling of tubes may result an incorrect additive-to-blood ratio, which could cause erroneous test results that could lead to improper diagnosis and/or treatment. BD also states that it has received no reports of serious injuries or deaths associated with this problem.

[View PDF](#)